

RADICALwaste

WEIGHBRIDGE & SCALES CALLIBRATION CERTIFICATES

WEIGHBRIDGE IN CERTIFICATE

Calibration Certificate Test Record

| CUSTOMER | MACHINE DETAILS | CERTIFICATE | TEST WEIGHTS |
|---|--|---|---|
| B & B Skip Hire Ltd. Beccles Business Park | <u>MAKE & MODEL</u> Lucid & Weighbridge | <u>CERT NO.</u> 578447H | <u>SET NO. / CERT NO.</u> Test Unit - EWT: TR11786/TN0035 |
| Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u> | <u>SERIAL NO.</u> 103719 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Inbound | <u>TESTED BY</u> Harry Scott <u>DATE TESTED</u> July 2023 <u>NEXT TEST DUE</u> July 2024 | |
| Julian Knights | <u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions | | |

AS FOUND

| LOAD POINT | READING | ERROR + / - |
|------------|---------|-------------|
| 0kg | 0kg | 0kg |
| 4000kg | 4000kg | 0kg |
| 8000kg | 8000kg | 0kg |
| 12000kg | 12000kg | 0kg |
| 16000kg | 16000kg | 0kg |
| 20000kg | 20000kg | 0kg |
| 30000kg | 30000kg | 0kg |
| 40000kg | 40000kg | 0kg |
| 50000kg | 50000kg | 0kg |
| | | |
| | | |
| | | |

DEFINITIVE

| LOAD POINT | READING | ERROR + / - |
|------------|---------|-------------|
| 0kg | 0kg | 0kg |
| 4000kg | 4000kg | 0kg |
| 8000kg | 8000kg | 0kg |
| 12000kg | 12000kg | 0kg |
| 16000kg | 16000kg | 0kg |
| 20000kg | 20000kg | 0kg |
| 30000kg | 30000kg | 0kg |
| 40000kg | 40000kg | 0kg |
| 50000kg | 50000kg | 0kg |
| | | |
| | | |
| | | |

CORNERING AS FOUND

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1 | 16000kg | 16000kg | 0kg |
| 2 | 16000kg | 16000kg | 0kg |
| 3 | 16000kg | 16000kg | 0kg |
| 4 | 16000kg | 16000kg | 0kg |
| | | | |
| | | | |
| | | | |
| | | | |

CORNERING DEFINITIVE

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1 | 16000kg | 16000kg | 0kg |
| 2 | 16000kg | 16000kg | 0kg |
| 3 | 16000kg | 16000kg | 0kg |
| 4 | 16000kg | 16000kg | 0kg |
| | | | |
| | | | |
| | | | |
| | | | |

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 

WEIGHBRIDGE OUT CERTIFICATE

Calibration Certificate Test Record

| CUSTOMER | MACHINE DETAILS | CERTIFICATE | TEST WEIGHTS |
|--|---|---|---|
| B & B Skip Hire Ltd. Beccles Business Park Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u> Julian Knights | <u>MAKE & MODEL</u> Lucid & Weighbridge <u>SERIAL NO.</u> 215676 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Outbound I.D TOLERANCE +/- 3 divisions | <u>CERT NO.</u> 578448H <u>TESTED BY</u> Harry Scott <u>DATE TESTED</u> July 2023 <u>NEXT TEST DUE</u> July 2024 | <u>SET NO. / CERT NO.</u> Test Unit - EWT: TR11786/TN0035 |

AS FOUND

| LOAD POINT | READING | ERROR + / - |
|------------|---------|-------------|
| 0kg | 0kg | 0kg |
| 6000kg | 6000kg | 0kg |
| 10000kg | 10000kg | 0kg |
| 16000kg | 16000kg | 0kg |
| 22000kg | 22000kg | 0kg |
| 30000kg | 30000kg | 0kg |
| 40000kg | 40000kg | 0kg |
| 50000kg | 50000kg | 0kg |
| | | |
| | | |
| | | |
| | | |

DEFINITIVE

| LOAD POINT | READING | ERROR + / - |
|------------|---------|-------------|
| 0kg | 0kg | 0kg |
| 6000kg | 6000kg | 0kg |
| 10000kg | 10000kg | 0kg |
| 16000kg | 16000kg | 0kg |
| 22000kg | 22000kg | 0kg |
| 30000kg | 30000kg | 0kg |
| 40000kg | 40000kg | 0kg |
| 50000kg | 50000kg | 0kg |
| | | |
| | | |
| | | |
| | | |

CORNERING AS FOUND

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1 | 10000kg | 10000kg | 0kg |
| 2 | 10000kg | 10000kg | 0kg |
| 3 | 10000kg | 10000kg | 0kg |
| 4 | 10000kg | 10000kg | 0kg |
| | | | |
| | | | |
| | | | |
| | | | |

CORNERING DEFINITIVE

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1 | 10000kg | 10000kg | 0kg |
| 2 | 10000kg | 10000kg | 0kg |
| 3 | 10000kg | 10000kg | 0kg |
| 4 | 10000kg | 10000kg | 0kg |
| | | | |
| | | | |
| | | | |
| | | | |

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 

SCALES CERTIFICATE

Calibration Certificate Test Record

| CUSTOMER | MACHINE DETAILS | CERTIFICATE | TEST WEIGHTS |
|---|--|---|---|
| B & B Skip Hire Ltd. Beccles Business Park | <u>MAKE & MODEL</u> A12 Platform Scale | <u>CERT NO.</u> 578449H | <u>SET NO. / CERT NO.</u> Test Unit - EWT: TR11786/TN0035 |
| Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u> | <u>SERIAL NO.</u> 81027002 <u>CAPACITY</u> 1500kg <u>DIVISION</u> 0.5kg <u>LOCATION</u> Bagging | <u>TESTED BY</u> Harry Scott <u>DATE TESTED</u> July 2023 <u>NEXT TEST DUE</u> July 2024 | |
| Julian Knights | <u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions | | |

AS FOUND

| LOAD POINT | READING | ERROR + / - |
|------------|----------|-------------|
| 0.0kg | 0.0kg | 0.0kg |
| 100.0kg | 100.0kg | 0.0kg |
| 200.0kg | 200.0kg | 0.0kg |
| 500.0kg | 500.0kg | 0.0kg |
| 1000.0kg | 1000.0kg | 0.0kg |
| 1500.0kg | 1500.0kg | 0.0kg |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |

DEFINITIVE

| LOAD POINT | READING | ERROR + / - |
|------------|----------|-------------|
| 0.0kg | 0.0kg | 0.0kg |
| 100.0kg | 100.0kg | 0.0kg |
| 200.0kg | 200.0kg | 0.0kg |
| 500.0kg | 500.0kg | 0.0kg |
| 1000.0kg | 1000.0kg | 0.0kg |
| 1500.0kg | 1500.0kg | 0.0kg |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |
| | | |

CORNERING AS FOUND

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1 | 200.0kg | 200.0kg | 0.0kg |
| 2 | 200.0kg | 200.0kg | 0.0kg |
| 3 | 200.0kg | 200.0kg | 0.0kg |
| 4 | 200.0kg | 200.0kg | 0.0kg |
| | | | |
| | | | |
| | | | |
| | | | |

CORNERING DEFINITIVE

| CORNER No. | LOAD POINT | READING | ERROR + / - |
|------------|------------|---------|-------------|
| 1 | 200.0kg | 200.0kg | 0.0kg |
| 2 | 200.0kg | 200.0kg | 0.0kg |
| 3 | 200.0kg | 200.0kg | 0.0kg |
| 4 | 200.0kg | 200.0kg | 0.0kg |
| | | | |
| | | | |
| | | | |
| | | | |

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 