

RADICALwaste

WEIGHBRIDGE & SCALES CALLIBRATION CERTIFICATES

WEIGHBRIDGE IN CERTIFICATE

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd. Beccles Business Park	<u>MAKE & MODEL</u> Lucid & Weighbridge	<u>CERT NO.</u> 581641J	<u>SET NO. / CERT NO.</u> Test Unit - EWT: TR12274/TP0057
Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u>	<u>SERIAL NO.</u> 103719 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Inbound	<u>TESTED BY</u> Scott Mitchell <u>DATE TESTED</u> July 2024 <u>NEXT TEST DUE</u> July 2025	
Julian Knights	<u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions		

AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
4000kg	4000kg	0kg
8000kg	8000kg	0kg
12000kg	12000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
4000kg	4000kg	0kg
8000kg	8000kg	0kg
12000kg	12000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 

WEIGHBRIDGE OUT CERTIFICATE



Unit 1, Cygnus, Orion Court,
Great Blakenham, Ipswich, Suffolk,
IP6 0LW
Tel: 01473 833360
info@stcsolutions.co.uk

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd. Beccles Business Park	<u>MAKE & MODEL</u> Lucid & Weighbridge	<u>CERT NO.</u> 581643J	<u>SET NO. / CERT NO.</u> Test Unit - EWT: TR12274/TP0057
Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u>	<u>SERIAL NO.</u> 215676 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Outbound	<u>TESTED BY</u> Scott Mitchell <u>DATE TESTED</u> July 2024 <u>NEXT TEST DUE</u> July 2025	
Julian Knights	<u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions		

AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
6000kg	6000kg	0kg
10000kg	10000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
6000kg	6000kg	0kg
10000kg	10000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 

SCALES CERTIFICATE

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd. Beccles Business Park	<u>MAKE & MODEL</u> A12 Platform Scale	<u>CERT NO.</u> 581642J	<u>SET NO. / CERT NO.</u> Test Unit - EWT: TR12274/TP0057
Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u>	<u>SERIAL NO.</u> 81027002 <u>CAPACITY</u> 1500kg <u>DIVISION</u> 0.5kg <u>LOCATION</u> Bagging	<u>TESTED BY</u> Scott Mitchell <u>DATE TESTED</u> July 2024 <u>NEXT TEST DUE</u> July 2025	
Julian Knights	<u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions		

AS FOUND

LOAD POINT	READING	ERROR + / -
0.0kg	0.0kg	0.0kg
100.0kg	100.0kg	0.0kg
200.0kg	200.0kg	0.0kg
500.0kg	500.0kg	0.0kg
1000.0kg	999.5kg	- 0.5kg
1500.0kg	1499.0kg	- 1.0kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0.0kg	0.0kg	0.0kg
100.0kg	100.0kg	0.0kg
200.0kg	200.0kg	0.0kg
500.0kg	500.0kg	0.0kg
1000.0kg	1000.0kg	0.0kg
1500.0kg	1500.0kg	0.0kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 