

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd. Beccles Business Park	<u>MAKE & MODEL</u> Lucid & Weighbridge	<u>CERT NO.</u> 584557H	<u>SET NO. / CERT NO.</u> Test Unit: EWT Weights TR12695/98182
Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u>	<u>SERIAL NO.</u> 215676 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Outbound	<u>TESTED BY</u> Scott Mitchell <u>DATE TESTED</u> July 2025 <u>NEXT TEST DUE</u> July 2026	
Julian Knights	<u>I.D.</u> <u>TOLERANCE</u> +/- 3 divisions		

AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
6000kg	6000kg	0kg
10000kg	10000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
6000kg	6000kg	0kg
10000kg	10000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager

Signature.

