Unit 1, Cygnus, Orion Court,
Great Blakenham, Ipswich, Suffolk,
IP6 0LW
Tel: 01473 833360
info@stcsolutions.co.uk

Calibration Certificate
Test Record

CUSTOMER	MACHINE DETAILS		CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd.	MAKE & MODEL		CERT NO.	SET NO. / CERT NO.
Beccles Business Park	A12 Platform Scale		584559H	Test Unit: EWT Weights TR12695/98182
Anson Way	SERIAL NO.	81027002	TESTED BY	
Ellough	CAPACITY	1500kg	Scott Mitchell	
Beccles	DIVISION	0.5kg	DATE TESTED	
Suffolk NR34 7TL	LOCATION	Bagging	July 2025	
CONTACT			NEXT TEST DUE	
	I.D		July 2026	
Julian Knights	TOLERANCE	+/- 3 divisions		

AS FOUND

LOAD POINT	READING	ERROR + / -
REPAIRED MACHINE – NO AS FOUND		

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0.0kg	0.0kg	0.0kg
100.0kg	100.0kg	0.0kg
200.0kg	200.0kg	0.0kg
500.0kg	500.0kg	0.0kg
1000.0kg	1000.0kg	0.0kg
1500.0kg	1500.0kg	0.0kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position**. Service Manager **Signature**.