Unit 1, Cygnus, Orion Court,
Great Blakenham, Ipswich, Suffolk,
IP6 0LW
Tel: 01473 833360
info@stcsolutions.co.uk

Calibration Certificate
Test Record

CUSTOMER	MACHINE DETAILS		CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd.	MAKE & MODEL		CERT NO.	SET NO. / CERT NO.
Beccles Business Park	Lucid & Weighbridge		587637H	Test Unit: EWT Weights TR13194
Anson Way	SERIAL NO.	103719	TESTED BY	
Ellough	CAPACITY	50000kg	Scott Mitchell	
Beccles	DIVISION	20kg	DATE TESTED	
Suffolk NR34 7TL	LOCATION	Inbound	July 2026	
CONTACT			NEXT TEST DUE	
	I.D		July 2027	
Julian Knights	TOLERANCE	+/- 3 divisions		

AS FOUND

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
400kg	400kg	0kg
6000kg	6000kg	0kg
10000kg	10000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg
400kg	400kg	0kg
6000kg	6000kg	0kg
10000kg	10000kg	0kg
16000kg	16000kg	0kg
20000kg	20000kg	0kg
30000kg	30000kg	0kg
40000kg	40000kg	0kg
50000kg	50000kg	0kg

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
1	8000kg	8000kg	0kg
2	8000kg	8000kg	0kg
3	8000kg	8000kg	0kg
4	8000kg	8000kg	0kg

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
1	8000kg	8000kg	0kg
2	8000kg	8000kg	0kg
3	8000kg	8000kg	0kg
4	8000kg	8000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position**. Service Manager **Signature**.

